

## **Rishabh Testing and Calibration Laboratory**

Plot No. C-6, MIDC, Satpur, Nasik-422007, Ph. 0253-2202370/307/309 Web: lab.rishabh.co.in mail:praveen.chaplot@rishabh.co.in



## **CE Declaration of Conformity**

Manufacturer:

SHANGHAI VA INSTRUMENT CO.LTD.

Address:

Area AD, 4th floor, building 22, 258 Yinlong Road, Waigang Town,

Jiading District, Shanghai, PRC SHANGHAI - 201806

Product Name:

Digital Multimeter

Model/ Type:

VA25A,VA25B,VA26A,VA28A,VA28B,VA29B

Rishabh testing and calibration laboratory hereby declares that the above products is in conformity with the

relevant union harmonization legislation:-

Sr No.	Directive
1	Electromagnetic Compatibility Regulations 2016
2	Electrical Equipment (Safety) Regulations 2016
3	The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

and following harmonized standard and technical specification have been applied:-

Standard	Standard Description
IEC 61326-1	Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements
EN 61326-2-2	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-2: Particular requirements - Test configurations, operational conditions and performance criteria for portable testing, measuring and monitoring equipment used in low-voltage distribution systems
IEC 61010-1	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements
IEC 61010-2-33	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-033: Particular requirements for hand-held multimeters and other meters for domestic and professional use, capable of measuring mains voltage
IEC 61010-031	Safety requirements for electrical equipment for measurement, control and laboratory use - Part 031: Safety requirements for hand-held and hand-manipulated probe assemblies for electrical test and measurement

Authorized Signatory

Ravikiran Deshmukh Head (QA & TS)

Date, Place 15.03.2023, NASHIK